

ATC Associates

104 East 25 Street, New York, NY 10010

Phone: (212) 353-8280 Fax: (212) 353-8306



Attn: Mr. Michael Zouak
Airttek Environmental Corporation
39 West 38th Sreet, 12th Floor
New York, NY 10018

Received: 05/16/07 4:15:00 PM
ATC Group #: 25048
Analysis Date: 05/16/07

Fax: (212) 768-0759 **Phone:** (212) 768-0516

Project: Fiterman Hall, Project# 06-0200
30 West Broadway, New York, NY

Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed by EPA 40 CFR Part 763 Final Rule (AHERA)

Sample	Volume	Area Analyzed	Non Asb	Asbestos Type(s)	# Structures		Analytical Sensitivity (S/cc)	Asbestos Concentration		Notes
					0.5µ-5	>5µ		(S/mm ²)	(S/cc)	
01 25048-1	1940.00	0.0450	0	None Detected	0	0	0.0044	<22.24	<0.0044	
02 25048-2	1940.00	0.0450	0	None Detected	0	0	0.0044	<22.24	<0.0044	
03 25048-3	1940.00	0.0450	0	None Detected	0	0	0.0044	<22.24	<0.0044	
04 25048-4	1940.00	0.0450	0	None Detected	0	0	0.0044	<22.24	<0.0044	

MARK PEYSAKHOV

Analyzed by:

MILENA BONEZZI

Approved by:

Disclaimer: The laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This lab is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by ATC Associates Inc. This report may not be used to claim product endorsement by NVLAP or any other agency of the U.S. Government. This reports relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysis' accuracy and precision) is available upon request.

Accredited for NVLAP PLM/TEM #101187-0, NY State ELAP #10879

Confidentiality Notice:

The document(s) contained herein are confidential and privileged information, intended for the exclusive use of the individual or entity named above.

Liability Notice:

ATC Associates Inc. and its personnel shall not be liable for any misinformation provided to us by the client regarding these samples. This report relates only to samples submitted and analyzed.

Unless otherwise indicated, no blank corrections were performed.

The condition of all samples was acceptable upon receipt.

Unless otherwise indicated all QC results were in control.

The laboratory certifies that the test results meet all requirements of NELAC.

Supplement to test report batch # _____, Amendments: _____, Amendment Dates:

Attn: Mr. Michael Zouak
 Airtek Environmental Corporation
 39 West 38th Sreet, 12th Floor
 New York, NY 10018

Received: 05/16/07 4:15:00 PM
 ATC Group #: 25048
 Analysis Date: 05/16/07

Fax: (212) 768-0759 Phone: (212) 768-0516

Project: Fiterman Hall, Project# 06-0200
 30 West Broadway, New York, NY

**Asbestos Fiber Analysis by Transmission Electron Microscopy (TEM) Performed
 by EPA 40 CFR Part 763 Final Rule (AHERA)**

Sample	Volume	Area Analyzed	Non Asb	Asbestos Type(s)	# Structures		Analytical Sensitivity (S/cc)	Asbestos Concentration		Notes
					0.5µ-5	>5µ		(S/mm ²)	(S/cc)	
05 25048-5	1940.00	0.0450	0	None Detected	0	0	0.0044	<22.24	<0.0044	
06 25048-6	1940.00	0.0450	0	None Detected	0	0	0.0044	<22.24	<0.0044	
07 25048-7	1820.00	0.0450	0	None Detected	0	0	0.0047	<22.24	<0.0047	

MARK PEYSAKHOV

Analyzed by:



MILENA BONEZZI

Approved by:



Disclaimer: The laboratory is not responsible for data reported in structures/cc, which is dependent on volume collected by non-laboratory personnel. This lab is only responsible for data reported in structures/mm². This report may not be reproduced, except in full, without written approval by ATC Associates Inc. This report may not be used to claim product endorsement by NVLAP or any other agency of the U.S. Government. This reports relates only to the samples reported above. Quality control data (including 95% confidence limits and laboratory and analysis' accuracy and precision) is available upon request.

Accredited for NVLAP PLM/TEM #101187-0, NY State ELAP #10879

Confidentiality Notice:

The document(s) contained herein are confidential and privileged information, intended for the exclusive use of the individual or entity named above.

Liability Notice:

ATC Associates Inc. and its personnel shall not be liable for any misinformation provided to us by the client regarding these samples. This report relates only to samples submitted and analyzed.

Unless otherwise indicated, no blank corrections were performed.

The condition of all samples was acceptable upon receipt.

Unless otherwise indicated all QC results were in control.

The laboratory certifies that the test results meet all requirements of NELAC.

Supplement to test report batch # _____, Amendments: _____, Amendment Dates:

25048

ASBESTOS AIR SAMPLING CHAIN OF CUSTODY

AIRTEK ENVIRONMENTAL
39 W 38TH STREET, 12TH FLOOR
NEW YORK, NY 10018
PHONE: 212 768-0516 FAX: 212 768-0759

PROJECT INFORMATION (fill out ALL information)

CLIENT: DASNY		PROJECT NAME: Fiterman Hall		SAMPLING START DATE: 04/15/07	
CLIENT ADDRESS: One Penn Plaza, New York, NY		PROJECT ADDRESS: 30 West Broadway, New York, NY		PROJECT #: 07-0050	
Turnaround Time: 24 HRS	Method: TEM/ AHRA	Laboratory: ATC	PROJECT PHASE: Asbestos Project Monitoring	Comments: Please email results to: schen@airtekeny.com, hennlewis@airtekeny.com, and emackenzie@airtekeny.com	
			SAMPLING TECHNICIAN: <i>Erin Mackenzie</i>	LOT #: 1415	
			PROJECT MANAGER: Erin Mackenzie		
			ROTOMETER #: 3/05A		

Sample #	Sample Media	Sample Location	Total Minutes			Flow Rate (L/min)			Contaminant	Conc.	
			Start	Stop	Total	Leak? Y/N	Final	Leak? Y/N			Avg. Volume
FH-051507-01-A	MCE-25	In the Personal Area	0710	1515	485	N	04	N	04	1940	Asbestos
FH-051507-02-A	MCE-25	Outside the Personal Area	0711	1576	485	N	04	N	04	1940	Asbestos
FH-051507-03-A	MCE-25	Inside the waste Area	0712	1577	485	N	04	N	04	1940	Asbestos
FH-051507-04-A	MCE-25	Outside the waste Area	0713	1518	485	N	04	N	04	1940	Asbestos
FH-051507-05-A	MCE-25	Next to Miced Trap Exhaust	0714	1579	485	N	04	N	04	1940	Asbestos
FH-051507-06-A	MCE-25	Next to the Fence	0715	1580	485	N	04	N	04	1940	Asbestos
FH-051507-07-A	MCE-25	Next to Freight Elevator Exhaust	0725	1508	485	N	04	N	04	1820	Asbestos
FH-051507-08-A	MCE-25	Field Blank									Asbestos
FH-051507-09-A	MCE-25	Field Blank									Asbestos
FH- -A	MCE-25										Asbestos

LAB INFORMATION

CHAIN OF CUSTODY		LAB INFORMATION	
Relinquished By (Print/Sign) <i>Erin Mackenzie</i>	Date/Time 04/15/07	Received By (Print/Sign) <i>Joe</i>	Date/Time 5/15/07 16:15
		Analyzed By (Print/Sign) <i>[Signature]</i>	Date/Time 5/15/07 16:15